

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130185</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Christopher Umbs</u>	ID# <u>3669</u> Date <u>1/10/12</u>
A Agency <u>Alaska State Troopers</u>	Phone # <u>465-4600</u>
Instrument Location <u>Gustavus</u>	
Alco S/N <u>2K</u> Target Value <u>.080</u> High Pressure <u>700 psi</u>	
B	
Alco Test Values <u>.082</u> <u>.082</u>	
1st Alco 2nd Alco	
Signature <u>[Signature]</u>	
(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of Feb, 2012 YK

[Signature] (Notary Seal Stamp)  
~~Daniel L. Nicholas~~ Nikki Roth  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130185 BMB  
1/18/12

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130185

JANUARY 18, 2012 BMB  
1/18/12

OPERATOR'S NAME:

UNOS, CHRISTOPHER A J

OPERATOR'S NUMBER: 3553

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME:

VOC

O.L. #: VOC

DEPT/AGENCY: AST

ORIG/REPORT: VOC

TEST TYPE: V

ALCO F-SET VALUE: .330

ALCO S/N: 2K

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.51 in

ALCO TARGET .002 10:10

BLANK TEST .000 10:12

INTERNAL STANDARD VERIFIED 10:12

ALCO TV 30.51 in .002 10:12

BLANK TEST .000 10:13

SUBJECT SAMPLE .330 10:14

BLANK TEST .000 10:14

ALCO TV 30.51 in .002 10:14

BLANK TEST .000 10:15

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130185

JANUARY 18, 2012

TIME 10:16

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/30/09

HEATERS

SAMPLE CHAMBER: 43c

BREATH TUBE: 41c

ZARONETER: 30.81 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

10120406-1012040670110-70120406  
HTALVCPOSTVWVWZT-1-abcde/ghijklm  
pqrstuvwxyz(123)

